

Analysenzertifikat Certificate of Analysis

SERION KBR Antigen / CFT Antigen LEGIONELLA PNEUMOPHILA

Best.-Nr. / Order No.	1224
Chargen-Nr. / Lot No.	SMI.CZ
Verfallsdatum / Expiry Date	2028-12
Antigen / Antigen	Legionella pneumophila subsp. Pneumophila, strain Philadelphia-1
Kultursystem / Culture System	Agar-Plattenkultur / cultivation on agar plates
Präparation / Preparation	beschallter Ganzzellextrakt sonicated cell lysate
Inaktivierung / Inactivation	Zerstörung der Zellen durch Ultraschall / Disruption of cells by sonication
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:32



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de