



Analysenzertifikat Certificate of Analysis

virion\sersion

SERION KBR Antigen / CFT Antigen PARAINFLUENZA VIRUS 1

Best.-Nr. / Order No.	1116
Chargen-Nr. / Lot No.	SKF.BT
Verfallsdatum / Expiry Date	2025-10
Antigen / Antigen	Parainfluenza Virus Typ 1 (Stamm Sendai) / Parainfluenza virus type 1 (Strain Sendai)
Kultursystem / Culture System	MDCK Zellen / MDCK cells
Präparation / Preparation	Infizierte Zellen, partiell aufgereinigt / Infected cells, partially purified
Inaktivierung / Inactivation	Hitze-Inaktivierung bei 56°C / Heat treatment at 56°C
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10



Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de