



Analysenzertifikat Certificate of Analysis

virion\sersion

SERION KBR Kontrollantigen / CFT Control Antigen ADENOVIRUS

Best.-Nr. / Order No.	2121
Chargen-Nr. / Lot No.	SDD.CN
Verfallsdatum / Expiry Date	2023-04
Kultursystem / Culture System	Verozellen / vero cells
Präparation / Preparation	Konzentrierter Zellkulturüberstand und alkalisch extrahierte Zellen concentrated cell culture supernatant and cells extracted with alkaline buffer
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10

Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany

Tel. +49 931 3045 0

Fax +49 931 3045 100

E-Mail: dialog@virion-serion.de

Internet: www.virion-serion.de

serion



multianalyt™