

Analysenzertifikat

Certificate of Analysis

SERION KBR Kontrollantigen / CFT Control Antigen
CHLAMYDIA

Best.-Nr. / Order No.	2122
Chargen-Nr. / Lot No.	SCI.AH
Verfallsdatum / Expiry Date	2028-03
Kultursystem / Culture System	Verozellen / vero cells
Präparation / Preparation	Konzentrierter Zellkulturüberstand und alkalisch extrahierte Zellen concentrated cell culture supernatant and cells extracted with alkaline buffer
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: dialog@virion-serion.de
 Internet: www.virion-serion.de