



Institut Virion\Serion GmbH · Postfach 92 53 · 97092 Würzburg

# Analysenzertifikat Certificate of Analysis

**virion\serion**  
Tel. +49 (0) 931 3045 0  
Fax +49 (0) 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)

## SERION KBR Negativkontrolle / CFT Negative Control COXSACKIEVIRUS B4

Best.-Nr. / Order No.	4175
Chargen-Nr. / Lot No.	SAB.BG
Verfallsdatum / Expiry Date	2021-01
Material / Material	Fötales Kälberserum / Fetal calf serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	< 1:10

serion



multianalyt™

Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)