

Analysezertifikat

Certificate of Analysis

SERION KBR Negativkontrolle / CFT Negative Control

LISTERIA MONOCYTOGENES

| | |
|--|--|
| Best.-Nr. / Order No. | 4234 |
| Chargen-Nr. / Lot No. | SCK.DG |
| Verfallsdatum / Expiry Date | 2028-10 |
| Material / Material | Fötale Kälberserum / Fetal calf serum |
| Lagerung / Storage | 2 °C - 8 °C |
| Rekonstitution / Reconstitution | 0,1 ml Aqua dest. |
| Gebrauchsverdünnung / Working Dilution | 1:10 |
| KBR Titer / CFT Titre | < 1:10 |



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de