

# Analysezertifikat Certificate of Analysis

## SERION KBR Negativkontrolle / CFT Negative Control MASERN VIRUS

Best.-Nr. / Order No.	4190
Chargen-Nr. / Lot No.	SLI.AR
Verfallsdatum / Expiry Date	2028-10
Material / Material	Fötale Kälberserum / Fetal calf serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	< 1:10



Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de