

# Analysenzertifikat

## Certificate of Analysis

### SERION KBR Positivkontrolle / CFT Positive Control CHLAMYDIA

Best.-Nr. / Order No.	3122
Chargen-Nr. / Lot No.	KK0040
IFU-Version / IFU version	21-19/07-1
Verfallsdatum / Expiry Date	2029-08-31
Material / Material	Humanserum / human serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	1:40



Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de