

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* ADENOVIRUS IgM

Best.-Nr. / Order No.	ESR128M
Chargen-Nr. / Lot No.	EK0136
Verfallsdatum / Expiry Date	2021-10-31
IFU-Version / IFU-Version	NEW! 128-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECK0265	1,07	> 0,54
Cut off-Serum / Cut off serum	ECK0266	0,29	0,15 – 0,58

Negativkontrolle / Negative Control	ADENOVIRUS IgM
Chargen-Nr. / Lot No.	ECK0264
Teststreifen / Strips	ADENOVIRUS IgM
Chargen-Nr. / Lot No.	ECK0272
Konjugat / Conjugate	IgM ++
Chargen-Nr. / Lot No.	KJK010

22.11.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de